



Plant Protection Division  
Weights and Measures Program  
Metrology Laboratory  
2747 29th Ave. SW ♦ Tumwater ♦ Washington 98512  
Ph (360) 753-5042 ♦ Fax (360) 586-4728  
e-mail: dwright@agr.wa.gov

**NVLAP**<sup>®</sup>  
For Scope of Accreditation  
Under  
NVLAP Lab Code 200446-0

### Report of Calibration # L2669-1

Calibration Date: November 19, 2007

#### Artifact

Test Item.....:	Weight Cart	Work Order #...:	L2669-1
Serial Number...:	103-2-90	Manufacture....:	Weight Cart, Inc.
Specification.....:	NIST HB 105-8	Material.....:	Cast Iron
Date Received...:	November 13, 2007	Seal Number....:	0264701

#### Submitted By

Washington State Patrol  
8623 Armstrong Road SW  
Olympia, WA 98504

Purchase Order # : N/A  
Point of Contact.: Dave Cromer  
Phone #.....: (360) 596-6000

#### Artifact Calibration Results

Weight Description	CM Correction As Found (lb)	CM Correction As Left (lb)	Class Weight Cart Tolerance ± (lb)	Uncertainty k=2 (lb)
4000 lb	<b>-15.85</b>	-0.33	1.25	0.42

#### Calibration Notes

- The artifact(s) listed above have been found and/or left within the stated tolerances for the classification(s) stated above, except as noted. An artifact is considered in-tolerance when the correction plus the measurement uncertainty is equal to or less than the specified tolerance. **Bold** print indicates an out-of-tolerance reading.
- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as 'apparent mass', scale verses 8.0 g/cm<sup>3</sup> reference mass density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C.

#### Calibration Conditions

Technician.....: Dan Wright

Procedure: NIST HB 145, SOP 4

Condition of Artifact: Good

Temperature.....: 19.9 °C

Pressure: 760.2 mm Hg

Humidity: 38.7 % RH

#### Laboratory Standard(s) Used

Description	Serial Number	Report Number	Date Calibrated	Date Calibration Due
2000 lb Mass Standards	SP2 & SP3	L2669-3	11/16/2007	11/16/2009
2000 lb Mass Standards	SP7 & SP11	L2669-3	11/16/2007	11/16/2009



Plant Protection Division  
Weights and Measures Program  
Metrology Laboratory  
2747 29th Ave. SW ♦ Tumwater ♦ Washington 98512  
Ph (360) 753-5042 ♦ Fax (360) 586-4728  
e-mail: dwright@agr.wa.gov

**NVLAP**<sup>®</sup>  
For Scope of Accreditation  
Under  
NVLAP Lab Code 200446-0

### ***Report of Calibration # L2669-1***

#### **Relevant Information**

- Liquid levels, as stated on the attached Weight Cart Inspection Checklist, must be maintained as close to reference levels as possible during use.
- The attached Weight Cart Inspection Checklist is an integral component of this Report of Calibration and a copy must be maintained with the cart and reviewed prior to use.
- Any maintenance, repairs, replacement of parts, or damage to the weight cart or its components will likely result in an out-of-tolerance condition. Maintenance or replacement of components such as batteries, tires, filters, or other items listed on the attached Weight Cart Inspection Checklist will require calibration of the weight cart prior to subsequent use.
- The results listed in this report relate only to the artifacts described and extent of calibrations performed.

#### **Traceability Statement**

The item(s) listed above have been compared to the Standards of the State of Washington. The Standards of the State of Washington are traceable to the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The report of calibration number identified in the title of this report is the unique report number to be used in referencing measurement traceability for the artifact(s) identified in this report only.

#### **Uncertainty Statement**

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits, and other uncertainties associated with the particular artifact (i.e., reading meniscus, air buoyancy corrections, etc.). The combined standard uncertainty is multiplied by  $k$ , a coverage factor of 2, to give the expanded uncertainty (which defines an interval with an approximate 95 percent level of confidence). The expanded uncertainty presented in this report is consistent with NIST Technical Note 1297. Stated uncertainties are less than 1/3 of the applicable tolerances. Magnetic testing has not been performed and there are no components for the effects of magnetism in the uncertainty budget.

#### **Certification Statement**

Accredited by the National Institute of Standards and Technology (NIST) through the National Voluntary Laboratory Accreditation Program (NVLAP) for the specified scope of accreditation under lab code 200446-0. This laboratory meets the requirements of ISO/IEC 17025 and ANSI/NCSL Z540-1. This report may not be used to claim product endorsement by NVLAP or any other government agency, and may not be reproduced, except in full, without written approval from the laboratory.

I certify (or declare) under penalty of perjury under the laws of the State of Washington that the foregoing is true and correct.

Executed this 19<sup>th</sup> day of November at Tumwater, Thurston County, Washington


Dan Wright

Dan Wright, State Metrologist



Metrology Laboratory Form  
WAMRF-011 4/9/03

## Weight Cart Inspection Checklist

<u>COMPANY</u> <b>Washington State Patrol</b>		<u>INSPECTION DATE</u> <b>November 19, 2007</b>	<u>STATE TEST NO.</u> <b>L2669-1</b>
<u>ADDRESS</u> <b>8623 Armstrong Road SW Olympia, WA 98504</b>		<u>NOMINAL VALUE</u> <b>4000 lb</b>	<u>MODEL NUMBER</u> <b>WC-20k</b>
		<u>MANUFACTURER</u> <b>Weight Carts, Inc.</b>	<u>SERIAL NUMBER</u> <b>103-2-90</b>
<u>POINT OF CONTACT</u> <b>Dave Cromer</b>		<u>PHONE NUMBER</u> <b>360-596-6000</b>	
<u>Power:</u> Electric Battery <input checked="" type="checkbox"/> ❖ Electric Generator <input type="checkbox"/> ❖ Gasoline <input type="checkbox"/> ❖ Diesel <input type="checkbox"/>			
<u>Fluids:</u>	<input type="checkbox"/> Engine Oil		Reference Level:
	<input type="checkbox"/> Hydraulic Oil	Sealed? Yes <input type="checkbox"/> No <input type="checkbox"/>	Reference Level:
	<input checked="" type="checkbox"/> Battery	Sealed? Yes <input type="checkbox"/> No <input checked="" type="checkbox"/>	Reference Level: @ lip
	<input type="checkbox"/> Fuel	Sight Gauge? Yes <input type="checkbox"/> No <input type="checkbox"/>	Reference Level:
<u>Number of axles:</u> <b>2</b>			
<u>Number / Size of Tires:</u> <b>2 @ 15 X 6 X 11 1/4 and 2 @ 21 X 7 X 15</b>			
<u>Nominal mass of weight cart is suitably marked.</u> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>			
<u>Do fluid drain tubes extend beyond the body of the cart?</u> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A <input checked="" type="checkbox"/>			
<u>Sealed wheel bearings.</u> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>			
<u>Drain holes present in locations where water may accumulate.</u> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>			
<u>Weight restraint railing permanently fixed and solid.</u> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>			
<u>Adjusting cavity is accessible.</u> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> Approximate capacity: <b>20 lb</b>			
<u>Adjusting cavity sealed.</u> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>			
<u>Service brakes are functioning properly.</u> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>			
<u>Parking brakes are functioning properly.</u> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>			
<u>Remote control is functioning properly.</u> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A <input checked="" type="checkbox"/>			
<u>General condition at time of calibration (i.e., any accumulated dirt/debris, damage, loose parts, or evidence of tampering or unauthorized entry of seals):</u> <b>Good</b>			
<u>List and report any repair and/or maintenance performed (i.e., leaks repaired, parts replaced, wheels changed, welding performed, etc.) since the last calibration:</u> <b>Batteries Replaced</b>			
Authorized Signature: <u></u>			